



Certificate of Conformance

Certificate Number: EXMXJ2301X82

Dated Issued: 03/04/2019

The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.¹

Product Item Number:	X100457
Item Description:	ASSY, BOX OF 5 AMALGATOME MD PKG
Item Lot Number:	X-0000000000082
Expiration Date:	01/2021

The blades were manufactured using a validated sterilization process:

STERIS Gamma Process Run ID:	103167A		
Processing Run Date:	16-FEB-2019		
Minimum Specified Dose (kGy):	25.0	Minimum Delivered Dose (kGy):	27.9
Maximum Specified Dose (kGy):	45.0	Maximum Delivered Dose (kGy):	37.1

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company. Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE.**



 Quality Department

03-04-2019

 Date



 Quality Manager

03-04-2019

 Date

¹ Manufacturing data filed within the Quality Management System and available onsite.