



## Certificate of Conformance

Certificate Number: EXMJ2273X28

Dated Issued: 03/04/2019

The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.<sup>1</sup>

<b>Product Item Number:</b>	X100593
<b>Item Description:</b>	Pack, Amalgatome SK head, 5 pc
<b>Item Lot Number:</b>	X-0000000000028
<b>Expiration Date:</b>	02/2021


The blades were manufactured using a validated sterilization process:

<b>STERIS Gamma Process Run ID:</b>	103167A		
<b>Processing Run Date:</b>	16-FEB-2019		
<b>Minimum Specified Dose (kGy):</b>	25.0	<b>Minimum Delivered Dose (kGy):</b>	27.9
<b>Maximum Specified Dose (kGy):</b>	45.0	<b>Maximum Delivered Dose (kGy):</b>	37.1

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company. Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE**.

  
 \_\_\_\_\_  
 Quality Department

03-04-2019  
 \_\_\_\_\_  
 Date

  
 \_\_\_\_\_  
 Quality Manager

03-04-2019  
 \_\_\_\_\_  
 Date

<sup>1</sup> Manufacturing data filed within the Quality Management System and available onsite.