



Certificate of Conformance

Certificate Number: EXMJ2323X29

Dated Issued: 03/04/2019

The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.¹

Product Item Number:	X100593
Item Description:	Pack, Amalgatome SK head, 5 pc
Item Lot Number:	X-0000000000029
Expiration Date:	02/2021

The blades were manufactured using a validated sterilization process:

STERIS Gamma Process Run ID:	103237A		
Processing Run Date:	23-FEB-2019		
Minimum Specified Dose (kGy):	25.0	Minimum Delivered Dose (kGy):	31.3
Maximum Specified Dose (kGy):	45.0	Maximum Delivered Dose (kGy):	39.4

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company. Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE.**



 Quality Department

03-04-2019

 Date



 Quality Manager

03-05-2019

 Date

¹ Manufacturing data filed within the Quality Management System and available onsite.