



Certificate of Conformance

Certificate Number: EXMXJ2361X86

Dated Issued: 20-MAR-2019


The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.¹

Product Item Number:	X100457
Item Description:	ASSY, BOX OF 5 AMALGATOME MD PKG
Item Lot Number:	X-00000000000086
Expiration Date:	2021/03


The blades were manufactured using a validated sterilization process:

STERIS Gamma Process Run ID:	103547A		
Processing Run Date:	13-MAR-2019		
Minimum Specified Dose (kGy):	25.0	Minimum Delivered Dose (kGy):	28.6
Maximum Specified Dose (kGy):	45.0	Maximum Delivered Dose (kGy):	37.9

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company. Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE.**


Quality Department

03-20-2019
Date


Quality Manager

03-20-2019
Date

¹ Manufacturing data filed within the Quality Management System and available onsite.