



## Certificate of Conformance

Certificate Number: EXMJ2445X88

Dated Issued: 23-APR-2019

The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.<sup>1</sup>

<b>Product Item Number:</b>	X100457
<b>Item Description:</b>	ASSY, BOX OF 5 AMALGATOME MD PKG
<b>Item Lot Number:</b>	X-0000000000088
<b>Expiration Date:</b>	01-OCT-2019


The blades were manufactured using a validated sterilization process:

<b>STERIS Gamma Process Run ID:</b>	104136A		
<b>Processing Run Date:</b>	18-APR-2019		
<b>Minimum Specified Dose (kGy):</b>	25.0	<b>Minimum Delivered Dose (kGy):</b>	29.5
<b>Maximum Specified Dose (kGy):</b>	45.0	<b>Maximum Delivered Dose (kGy):</b>	39.4

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company. Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE**.

  
Quality Department

23 Apr. 2019  
Date

  
Quality Manager

24 Apr. 2019  
Date

<sup>1</sup> Manufacturing data filed within the Quality Management System and available onsite.