



## Certificate of Conformance

Certificate Number: EXMXJ2451X89

Dated Issued: 05-JUN-2019

The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.<sup>1</sup>

<b>Product Item Number:</b>	X100457
<b>Item Description:</b>	ASSY, BOX OF 5 AMALGATOME MD PKG
<b>Item Lot Number:</b>	X-0000000000089
<b>Expiration Date:</b>	01-OCT-2019


The blades were manufactured using a validated sterilization process:

<b>STERIS Gamma Process Run ID:</b>	104907A		
<b>Processing Run Date:</b>	01-JUN-2019		
<b>Minimum Specified Dose (kGy):</b>	25.0	<b>Minimum Delivered Dose (kGy):</b>	28.8
<b>Maximum Specified Dose (kGy):</b>	45.0	<b>Maximum Delivered Dose (kGy):</b>	39.2

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company.  
Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE.**

  
Quality Department

05-JUN-2019  
Date

  
Quality Manager

05-JUNE-2019  
Date

<sup>1</sup> Manufacturing data filed within the Quality Management System and available onsite.