



Certificate of Conformance

Certificate Number: EXMXJ2452X90

Dated Issued: 05-JUN-2019


The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.¹

Product Item Number:	X100457
Item Description:	ASSY, BOX OF 5 AMALGATOME MD PKG
Item Lot Number:	X-0000000000090
Expiration Date:	01-OCT-2019

The blades were manufactured using a validated sterilization process:

STERIS Gamma Process Run ID:	104907A		
Processing Run Date:	01-JUN-2019		
Minimum Specified Dose (kGy):	25.0	Minimum Delivered Dose (kGy):	28.8
Maximum Specified Dose (kGy):	45.0	Maximum Delivered Dose (kGy):	39.2

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company.
Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE**.


Quality Department

05 June 2019
Date


Quality Manager

06 June 2019
Date

¹ Manufacturing data filed within the Quality Management System and available onsite.