



Certificate of Conformance

Certificate Number: EXMXJ2486X91

Dated Issued: 07-JUL-2019

The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.¹

Product Item Number:	X100457
Item Description:	ASSY, BOX OF 5 AMALGATOME MD PKG
Item Lot Number:	X-00000000000091
Expiration Date:	05-2021

The blades were manufactured using a validated sterilization process:

STERIS Gamma Process Run ID:	105371A		
Processing Run Date:	28-JUN-2019		
Minimum Specified Dose (kGy):	25.0	Minimum Delivered Dose (kGy):	28.3
Maximum Specified Dose (kGy):	45.0	Maximum Delivered Dose (kGy):	37.8

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company. Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE**.



 Quality Department



 Date



 Quality Manager



 Date

¹ Manufacturing data filed within the Quality Management System and available onsite.