



Certificate of Conformance

Certificate Number: EXMXJ2626X36

Dated Issued: 21-AUG-2019

The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.¹

Product Item Number:	X100593
Item Description:	Pack, Amalgatome SK head, 5 pc
Item Lot Number:	X-0000000000036
Expiration Date:	2021-07

The blades were manufactured using a validated sterilization process:

STERIS Gamma Process Run ID:	106182A		
Processing Run Date:	19-AUG-2019		
Minimum Specified Dose (kGy):	25.0	Minimum Delivered Dose (kGy):	28.8
Maximum Specified Dose (kGy):	45.0	Maximum Delivered Dose (kGy):	35.4

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company. Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE**.



 Quality Department



 Quality Manager

21-Aug-2019
 Date

21-Aug-2019
 Date

¹ Manufacturing data filed within the Quality Management System and available onsite.