



## Certificate of Conformance

Certificate Number: EXMXJ2627X37

Dated Issued: 04 SEPT 2019


The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.<sup>1</sup>

<b>Product Item Number:</b>	X100593
<b>Item Description:</b>	Pack, Amalgatome SK head, 5 pc
<b>Item Lot Number:</b>	X-00000000000037
<b>Expiration Date:</b>	2021/07

The blades were manufactured using a validated sterilization process:

<b>STERIS Gamma Process Run ID:</b>	106422A		
<b>Processing Run Date:</b>	31-AUG-2019		
<b>Minimum Specified Dose (kGy):</b>	25.0	<b>Minimum Delivered Dose (kGy):</b>	29.6
<b>Maximum Specified Dose (kGy):</b>	45.0	<b>Maximum Delivered Dose (kGy):</b>	38.4

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company. Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE.**

  
 \_\_\_\_\_  
 Quality Department

04-Sept 2019  
 Date

  
 \_\_\_\_\_  
 Quality Manager

24-Sept-2019  
 Date

<sup>1</sup> Manufacturing data filed within the Quality Management System and available onsite.