



Certificate of Conformance

Certificate Number: EXMXJ3018X58

Dated Issued: 04/09/2020

The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.¹

Product Item Number:	X100593
Item Description:	Pack, Amalgatome SK head, 5 pc
Item Lot Number:	X-00000000000058
Expiration Date:	2022/03

The blades were manufactured using a validated sterilization process:

STERIS Gamma Process Run ID:	109640A		
Processing Run Date:	01-Apr-2020		
Minimum Specified Dose (kGy):	25.0	Minimum Delivered Dose (kGy):	28.1
Maximum Specified Dose (kGy):	45.0	Maximum Delivered Dose (kGy):	38.5

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company. Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE**.



 Quality Department


 Quality Manager

04-09-2020

 Date

04-10-2020

 Date

¹ Manufacturing data filed within the Quality Management System and available onsite.