



## Certificate of Conformance

Certificate Number: EXMJ3070X60

Dated Issued: 06/08/2020

The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.<sup>1</sup>

<b>Product Item Number:</b>	X100593
<b>Item Description:</b>	Pack, Amalgatome SK head, 5 pc
<b>Item Lot Number:</b>	X-0000000000060
<b>Expiration Date:</b>	2022/04


The blades were manufactured using a validated sterilization process:

<b>STERIS Gamma Process Run ID:</b>	110266A		
<b>Processing Run Date:</b>	22-May-2020		
<b>Minimum Specified Dose (kGy):</b>	25.0	<b>Minimum Delivered Dose (kGy):</b>	26.9
<b>Maximum Specified Dose (kGy):</b>	45.0	<b>Maximum Delivered Dose (kGy):</b>	37.4

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company.  
Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE.**

  
Quality Department

  
Quality Manager  
for Kayli Camp

  
Date

06-08-2020  
Date

<sup>1</sup> Manufacturing data filed within the Quality Management System and available onsite.