



Certificate of Conformance

Certificate Number: EXMXJ3144X107

Dated Issued: 07/31/2020


The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.¹

Product Item Number:	X100457
Item Description:	ASSY, BOX OF 5 AMALGATOME MD PKG
Item Lot Number:	X-0000000000107
Expiration Date:	2022/07

The blades were manufactured using a validated sterilization process:

STERIS Gamma Process Run ID:	111036A		
Processing Run Date:	25-Jul-2020		
Minimum Specified Dose (kGy):	25.0	Minimum Delivered Dose (kGy):	26.8
Maximum Specified Dose (kGy):	45.0	Maximum Delivered Dose (kGy):	38.1

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company. Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE**.



 Quality Department


 Quality Manager

07/31/2020

 Date
 08-03-2020

 Date

¹ Manufacturing data filed within the Quality Management System and available onsite.