



Certificate of Conformance

Certificate Number: EXMXJ3190X108

Dated Issued: 09/03/2020

The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.¹


Product Item Number:	X100457
Item Description:	ASSY, BOX OF 5 AMALGATOME MD PKG
Item Lot Number:	X-0000000000108
Expiration Date:	2022/07

The blades were manufactured using a validated sterilization process:

STERIS Gamma Process Run ID:	111422A		
Processing Run Date:	28-Aug-2020		
Minimum Specified Dose (kGy):	25.0	Minimum Delivered Dose (kGy):	28.9
Maximum Specified Dose (kGy):	45.0	Maximum Delivered Dose (kGy):	38.7

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company. Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE**.



 Quality Department


 Quality Manager

09/03/2020
 Date

09-03-2020
 Date

¹ Manufacturing data filed within the Quality Management System and available onsite.