



Certificate of Conformance

Certificate Number: EXMXJ3396X115

Dated Issued: 03/18/2021

The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.¹

Product Item Number:	X100457
Item Description:	ASSY, BOX OF 5 AMALGATOME MD PKG
Item Lot Number:	X-0000000000115
Expiration Date:	2023/02


The blades were manufactured using a validated sterilization process:

STERIS Gamma Process Run ID:	1073-2594A		
Processing Run Date:	08-Mar-2021		
Minimum Specified Dose (kGy):	25.0	Minimum Delivered Dose (kGy):	27.5
Maximum Specified Dose (kGy):	45.0	Maximum Delivered Dose (kGy):	39.3

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company. Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE.**



 Quality Department



 Quality Manager
for Kayli Camp

3/18/21

 Date

03-18-21

 Date

¹ Manufacturing data filed within the Quality Management System and available onsite.