



Certificate of Conformance

Certificate Number: EXMXJ3397X73

Dated Issued: 03/16/2021

The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.¹

Product Item Number:	X100593
Item Description:	Pack, Amalgatome SK head, 5 pc
Item Lot Number:	X-0000000000073
Expiration Date:	2023/02

The blades were manufactured using a validated sterilization process:

STERIS Gamma Process Run ID:	1073-2594A		
Processing Run Date:	08-Mar-2021		
Minimum Specified Dose (kGy):	25.0	Minimum Delivered Dose (kGy):	27.5
Maximum Specified Dose (kGy):	45.0	Maximum Delivered Dose (kGy):	39.3

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company. Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE.**



 Quality Department

3/16/21

 Date

 Quality Manager

 Date

¹ Manufacturing data filed within the Quality Management System and available onsite.