



Certificate of Conformance

Certificate Number: EXMXJ3408X78

Dated Issued: 04/19/2021


The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.¹

Product Item Number:	X100593
Item Description:	Pack, Amalgatome SK Head, 5 PC
Item Lot Number:	X-0000000000078
Expiration Date:	2023/03

The blades were manufactured using a validated sterilization process:

STERIS Gamma Process Run ID:	1073-3275A		
Processing Run Date:	15-Apr-21		
Minimum Specified Dose (kGy):	25.0	Minimum Delivered Dose (kGy):	27.7
Maximum Specified Dose (kGy):	45.0	Maximum Delivered Dose (kGy):	39.3

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company. Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE.**


Quality Department

04/21/2021
Date


Quality Manager

04-21-2021
Date

¹ Manufacturing data filed within the Quality Management System and available onsite.