



Certificate of Conformance

Certificate Number: EXMXJ3447X118

Dated Issued: 05/11/2021

The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.¹

| | |
|-----------------------------|-----------------------------------|
| Product Item Number: | X100457 |
| Item Description: | Blade, Box of 5 Amalgatome MD PKG |
| Item Lot Number: | X-0000000000118 |
| Expiration Date: | 2023/04 |

The blades were manufactured using a validated sterilization process:

| | | | |
|--------------------------------------|------------|--------------------------------------|------|
| STERIS Gamma Process Run ID: | 1073-3685A | | |
| Processing Run Date: | 10-May-21 | | |
| Minimum Specified Dose (kGy): | 25.0 | Minimum Delivered Dose (kGy): | 27.2 |
| Maximum Specified Dose (kGy): | 45.0 | Maximum Delivered Dose (kGy): | 39.8 |


Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company. Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE**.



 Quality Department

05/13/2021

 Date



 Quality Manager

05/11/2021

 Date

¹ Manufacturing data filed within the Quality Management System and available onsite.