



Certificate of Conformance

Certificate Number: EXMXJ3461X119

Dated Issued: 06/15/2021



The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.¹

Product Item Number:	X100457
Item Description:	ASSY, BOX OF 5 AMALGATOME MD PKG
Item Lot Number:	X-0000000000119
Expiration Date:	2023/04

The blades were manufactured using a validated sterilization process:

STERIS Gamma Process Run ID:	1073-4187A		
Processing Run Date:	14-Jun-21		
Minimum Specified Dose (kGy):	25.0	Minimum Delivered Dose (kGy):	27.0
Maximum Specified Dose (kGy):	45.0	Maximum Delivered Dose (kGy):	39.4

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company. Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE**.


 Quality Department

 Quality Manager

06/15/2021
 Date

06-21-2021
 Date

¹ Manufacturing data filed within the Quality Management System and available onsite.