



## Certificate of Conformance

Certificate Number: EXMXJ3567X121

Dated Issued: 07/19/2021

The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.<sup>1</sup>

<b>Product Item Number:</b>	X100457
<b>Item Description:</b>	Blade, Box of 5 Amalgatome MD PKG
<b>Item Lot Number:</b>	X-0000000000121
<b>Expiration Date:</b>	2023/06


The blades were manufactured using a validated sterilization process:

<b>STERIS Gamma Process Run ID:</b>	1073-4683A		
<b>Processing Run Date:</b>	15-Jul-21		
<b>Minimum Specified Dose (kGy):</b>	25.0	<b>Minimum Delivered Dose (kGy):</b>	26.2
<b>Maximum Specified Dose (kGy):</b>	45.0	<b>Maximum Delivered Dose (kGy):</b>	38.5

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company. Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE**.

  
 \_\_\_\_\_  
 Quality Department

07/19/2021  
 \_\_\_\_\_  
 Date

  
 \_\_\_\_\_  
 Quality Manager  
 Joe Hayli Camp

07-23-2021  
 \_\_\_\_\_  
 Date

<sup>1</sup> Manufacturing data filed within the Quality Management System and available onsite.