



Certificate of Conformance

Certificate Number: EXMXJ3568X86

Dated Issued: 07/19/2021

The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.¹

Product Item Number:	X100593
Item Description:	Pack, Amalgatome SK Head, 5 PC
Item Lot Number:	X-0000000000086
Expiration Date:	2023-06

The blades were manufactured using a validated sterilization process:

STERIS Gamma Process Run ID:	1073-4683A		
Processing Run Date:	15-Jul-21		
Minimum Specified Dose (kGy):	25.0	Minimum Delivered Dose (kGy):	26.2
Maximum Specified Dose (kGy):	45.0	Maximum Delivered Dose (kGy):	38.5

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company. Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE.**



 Quality Department

07/19/2021

 Date



 Quality Manager

07-30-21

 Date

¹ Manufacturing data filed within the Quality Management System and available onsite.