



## Certificate of Conformance

Certificate Number: EXMXJ3603X122

Dated Issued: 08/30/2021

The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.<sup>1</sup>

<b>Product Item Number:</b>	X100457
<b>Item Description:</b>	BLADE, BOX OF 5 AMALGATOME MD PKG
<b>Item Lot Number:</b>	X-0000000000122
<b>Expiration Date:</b>	2023/07


The blades were manufactured using a validated sterilization process:

<b>STERIS Gamma Process Run ID:</b>	1073-5293A		
<b>Processing Run Date:</b>	26-Aug-2021		
<b>Minimum Specified Dose (kGy):</b>	25.0	<b>Minimum Delivered Dose (kGy):</b>	28.3
<b>Maximum Specified Dose (kGy):</b>	45.0	<b>Maximum Delivered Dose (kGy):</b>	40.0

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company. Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE**.

  
 \_\_\_\_\_  
 Quality Department

  
 \_\_\_\_\_  
 Quality Manager

  
 \_\_\_\_\_

08/30/2021  
 Date

08-30-2021  
 Date

<sup>1</sup> Manufacturing data filed within the Quality Management System and available onsite.