



Certificate of Conformance

Certificate Number: EXMXJ3624X95

Dated Issued: 10/18/2021

The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.¹

Product Item Number:	X100593
Item Description:	PACK, AMALGATOME SK HEAD, 5 PC
Item Lot Number:	X-0000000000095
Expiration Date:	2023-08

The blades were manufactured using a validated sterilization process:

STERIS Gamma Process Run ID:	1073-5856A		
Processing Run Date:	08-Oct-2021		
Minimum Specified Dose (kGy):	25.0	Minimum Delivered Dose (kGy):	27.6
Maximum Specified Dose (kGy):	45.0	Maximum Delivered Dose (kGy):	39.6


Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company. Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE.**



 Quality Department

10/18/21

 Date



 Quality Manager
 Tech Ops Supervisor

10/18/21

 Date

¹ Manufacturing data filed within the Quality Management System and available onsite.