



Certificate of Conformance

Certificate Number: EXMJ3673X124

Dated Issued: 11/01/2021

The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.¹

| | |
|-----------------------------|-----------------------------------|
| Product Item Number: | X100457 |
| Item Description: | BLADE, BOX OF 5 AMALGATOME MD PKG |
| Item Lot Number: | X-0000000000124 |
| Expiration Date: | 2023/08 |

The blades were manufactured using a validated sterilization process:

| | | | |
|--------------------------------------|-------------|--------------------------------------|------|
| STERIS Gamma Process Run ID: | 1073-5856A | | |
| Processing Run Date: | 08-Oct-2021 | | |
| Minimum Specified Dose (kGy): | 25.0 | Minimum Delivered Dose (kGy): | 27.6 |
| Maximum Specified Dose (kGy): | 45.0 | Maximum Delivered Dose (kGy): | 39.6 |

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company. Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE.**



 Quality Department



 Quality Manager

11/01/2021
 Date

11/01/21
 Date

¹ Manufacturing data filed within the Quality Management System and available onsite.