



Certificate of Conformance

Certificate Number: EXMJ3677X97

Dated Issued: 10/21/2021

The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.¹

Product Item Number:	X100593
Item Description:	PACK, AMALGATOME SK HEAD, 5 PC
Item Lot Number:	X-00000000000097
Expiration Date:	2023-09

The blades were manufactured using a validated sterilization process:

STERIS Gamma Process Run ID:	1073-5856A		
Processing Run Date:	08-Oct-2021		
Minimum Specified Dose (kGy):	25.0	Minimum Delivered Dose (kGy):	27.6
Maximum Specified Dose (kGy):	45.0	Maximum Delivered Dose (kGy):	39.6

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company. Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE**.



 Quality Department



 Quality Manager

10/21/2021
 Date

10/21/21
 Date

¹ Manufacturing data filed within the Quality Management System and available onsite.