



Certificate of Conformance

Certificate Number: EXMXJ3710X06

Dated Issued: 12/02/2021

The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.¹

Product Item Number:	X101835
Item Description:	ASK, CUTTING HEAD, 10 PC
Item Lot Number:	X-00000000000006
Expiration Date:	2023-10

The blades were manufactured using a validated sterilization process:

STERIS Gamma Process Run ID:	1073-6574A		
Processing Run Date:	20-Nov-2021		
Minimum Specified Dose (kGy):	25.0	Minimum Delivered Dose (kGy):	28.5
Maximum Specified Dose (kGy):	45.0	Maximum Delivered Dose (kGy):	39.0

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company. Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE**.


Quality Department

12/02/2021
Date


Quality Manager

12/3/21
Date

¹ Manufacturing data filed within the Quality Management System and available onsite.