



Certificate of Conformance

Certificate Number: EXMXJ3730X07

Dated Issued: 12/17/2021

The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.¹

Product Item Number:	X101835
Item Description:	ASK, CUTTING HEAD, 10 PC
Item Lot Number:	X-0000000000007
Expiration Date:	2023-11

The blades were manufactured using a validated sterilization process:

STERIS Gamma Process Run ID:	1073-6962A		
Processing Run Date:	13-Dec-2021		
Minimum Specified Dose (kGy):	25.0	Minimum Delivered Dose (kGy):	28.4
Maximum Specified Dose (kGy):	45.0	Maximum Delivered Dose (kGy):	40.5

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company. Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE.**


Quality Department

12/17/2021
Date


Quality Manager

12/21/21
Date

¹ Manufacturing data filed within the Quality Management System and available onsite.