



Certificate of Conformance

Certificate Number: EXMXJ3797X100

Dated Issued: 01/25/2022

The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.¹

Product Item Number:	X100593
Item Description:	PACK, AMALGATOME SK HEAD, 5 PC
Item Lot Number:	X-0000000000100
Expiration Date:	2023-12

The blades were manufactured using a validated sterilization process:

STERIS Gamma Process Run ID:	1073-7475A		
Processing Run Date:	24-Jan-2022		
Minimum Specified Dose (kGy):	25.0	Minimum Delivered Dose (kGy):	28.5
Maximum Specified Dose (kGy):	45.0	Maximum Delivered Dose (kGy):	40.1

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company. Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE.**



 Quality Department

01/25/2022

 Date



 Quality Manager

1/28/22

 Date

¹ Manufacturing data filed within the Quality Management System and available onsite.