



Certificate of Conformance

Certificate Number: EXMXJ3830X13

Dated Issued: 02/28/2022

The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.¹

Product Item Number:	X101835
Item Description:	ASK, CUTTING HEAD, 10 PC
Item Lot Number:	X-0000000000013
Expiration Date:	2024-01

The blades were manufactured using a validated sterilization process:

STERIS Gamma Process Run ID:	1073-7996A		
Processing Run Date:	25-Feb-2022		
Minimum Specified Dose (kGy):	25.0	Minimum Delivered Dose (kGy):	27.7
Maximum Specified Dose (kGy):	45.0	Maximum Delivered Dose (kGy):	38.3

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company. Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE**.



 Quality Department



 Quality Manager

02/28/2022

 Date

3/11/22

 Date

¹ Manufacturing data filed within the Quality Management System and available onsite.