



Certificate of Conformance

Certificate Number: EXMJ3859X102

Dated Issued: 03/28/2022

The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.¹

Product Item Number:	X100593
Item Description:	PACK, AMALGATOME SK HEAD, 5 PC
Item Lot Number:	X-0000000000102
Expiration Date:	2024-02

The blades were manufactured using a validated sterilization process:

STERIS Gamma Process Run ID:	1073-8412A		
Processing Run Date:	25-Mar-2022		
Minimum Specified Dose (kGy):	25.0	Minimum Delivered Dose (kGy):	27.8
Maximum Specified Dose (kGy):	45.0	Maximum Delivered Dose (kGy):	38.1

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company. Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE**.



 Quality Department


 Quality Manager

03/28/2022

 Date

03-28-2022

 Date

¹ Manufacturing data filed within the Quality Management System and available onsite.