



Certificate of Conformance

Certificate Number: EXMXJ3862X103

Dated Issued: 04/07/2022

The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.¹

Product Item Number:	X100593
Item Description:	PACK, AMALGATOME SK HEAD, 5 PC
Item Lot Number:	X-00000000000103
Expiration Date:	2024-02

The blades were manufactured using a validated sterilization process:

STERIS Gamma Process Run ID:	1073-8412A		
Processing Run Date:	25-Mar-2022		
Minimum Specified Dose (kGy):	25.0	Minimum Delivered Dose (kGy):	27.8
Maximum Specified Dose (kGy):	45.0	Maximum Delivered Dose (kGy):	38.1

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company. Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE.**



 Quality Department



 Quality Manager

04/07/2022

 Date

4/8/22

 Date

¹ Manufacturing data filed within the Quality Management System and available onsite.