



Certificate of Conformance

Certificate Number: EXMXJ3904X131

Dated Issued: 05/06/2022

The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.¹

Product Item Number:	X100457
Item Description:	BLADE, BOX OF 5 AMALGATOME MD PKG
Item Lot Number:	X-00000000000131
Expiration Date:	2024/03

The blades were manufactured using a validated sterilization process:

STERIS Gamma Process Run ID:	1073-8921A		
Processing Run Date:	29-Apr-2022		
Minimum Specified Dose (kGy):	25.0	Minimum Delivered Dose (kGy):	28.8
Maximum Specified Dose (kGy):	45.0	Maximum Delivered Dose (kGy):	40.1

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company. Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE.**


Quality Department

05/06/2022
Date


Quality Manager

5/6/22
Date

¹ Manufacturing data filed within the Quality Management System and available onsite.