



## Certificate of Conformance

Certificate Number: EXMXJ3936X106

Dated Issued: 06/01/2022

The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.<sup>1</sup>

<b>Product Item Number:</b>	X100593
<b>Item Description:</b>	PACK, AMALGATOME SK HEAD, 5 PC
<b>Item Lot Number:</b>	X-0000000000106
<b>Expiration Date:</b>	2024-05

The blades were manufactured using a validated sterilization process:

<b>STERIS Gamma Process Run ID:</b>	1073-9396A		
<b>Processing Run Date:</b>	30-May-2022		
<b>Minimum Specified Dose (kGy):</b>	25.0	<b>Minimum Delivered Dose (kGy):</b>	27.2
<b>Maximum Specified Dose (kGy):</b>	45.0	<b>Maximum Delivered Dose (kGy):</b>	37.4

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company. Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE**.

  
Quality Department

  
Quality Manager

06/01/2022  
Date

6/13/22  
Date

<sup>1</sup> Manufacturing data filed within the Quality Management System and available onsite.