



Certificate of Conformance

Certificate Number: EXMXJ3938X107

Dated Issued: 06/01/2022

The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.¹

Product Item Number:	X100593
Item Description:	PACK, AMALGATOME SK HEAD, 5 PC
Item Lot Number:	X-0000000000107
Expiration Date:	2024-05

The blades were manufactured using a validated sterilization process:

STERIS Gamma Process Run ID:	1073-9396A		
Processing Run Date:	30-May-2022		
Minimum Specified Dose (kGy):	25.0	Minimum Delivered Dose (kGy):	27.2
Maximum Specified Dose (kGy):	45.0	Maximum Delivered Dose (kGy):	37.4

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company. Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE.**



 Quality Department



 Quality Manager

06/01/2022
 Date

6/17/22
 Date

¹ Manufacturing data filed within the Quality Management System and available onsite.