



Certificate of Conformance

Certificate Number: EXMJ4128X137

Dated Issued: 10/26/2022

The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.¹

Product Item Number:	X100457
Item Description:	BLADE, BOX OF 5 AMALGATOME MD PKG
Item Lot Number:	X-0000000000137
Expiration Date:	2024/08

The blades were manufactured using a validated sterilization process:

STERIS Gamma Process Run ID:	1073-11453A		
Processing Run Date:	24-Oct-2022		
Minimum Specified Dose (kGy):	25.0	Minimum Delivered Dose (kGy):	27.6
Maximum Specified Dose (kGy):	45.0	Maximum Delivered Dose (kGy):	39.4

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company. Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE**.


Quality Department

10/26/2022
Date


Quality Manager

10/26/22
Date

¹ Manufacturing data filed within the Quality Management System and available onsite.