



Certificate of Conformance

Certificate Number: EXMXJ4189X26

Dated Issued: 12/05/2022

The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.¹

| | |
|-----------------------------|--------------------------|
| Product Item Number: | X101835 |
| Item Description: | ASK, CUTTING HEAD, 10 PC |
| Item Lot Number: | X-0000000000026 |
| Expiration Date: | 2024-11 |

The blades were manufactured using a validated sterilization process:

| | | | |
|--------------------------------------|-------------|--------------------------------------|------|
| STERIS Gamma Process Run ID: | 1073-11874A | | |
| Processing Run Date: | 21-Nov-2022 | | |
| Minimum Specified Dose (kGy): | 25.0 | Minimum Delivered Dose (kGy): | 27.7 |
| Maximum Specified Dose (kGy): | 45.0 | Maximum Delivered Dose (kGy): | 38.8 |

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company. Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE**.



 Quality Department

12/05/2022

 Date



 Quality Manager

12/5/22

 Date

¹ Manufacturing data filed within the Quality Management System and available onsite.