



Certificate of Conformance

Certificate Number: EXMXJ4201X28

Dated Issued: 01/06/2023

The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.¹

Product Item Number:	X101835
Item Description:	ASK, CUTTING HEAD, 10 PC
Item Lot Number:	X-0000000000028
Expiration Date:	2024-12

The blades were manufactured using a validated sterilization process:

STERIS Gamma Process Run ID:	1073-12322A		
Processing Run Date:	02-Jan-2023		
Minimum Specified Dose (kGy):	25.0	Minimum Delivered Dose (kGy):	28.6
Maximum Specified Dose (kGy):	45.0	Maximum Delivered Dose (kGy):	38.4

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company. Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE**.



 Quality Department



 Quality Manager

01/06/2023
 Date

1/9/23
 Date

¹ Manufacturing data filed within the Quality Management System and available onsite.