



Certificate of Conformance

Certificate Number: EXMXJ4282X141

Dated Issued: 02/23/2023

The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.¹

Product Item Number:	X100457
Item Description:	BLADE, BOX OF 5 AMALGATOME MD PKG
Item Lot Number:	X-0000000000141
Expiration Date:	2025/01

The blades were manufactured using a validated sterilization process:

STERIS Gamma Process Run ID:	1073-12962A		
Processing Run Date:	20-Feb-2023		
Minimum Specified Dose (kGy):	25.0	Minimum Delivered Dose (kGy):	26.8
Maximum Specified Dose (kGy):	45.0	Maximum Delivered Dose (kGy):	37.9

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company. Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE**.



 Quality Department



 Quality Manager

02/23/2023

 Date

2/23/23

 Date

¹ Manufacturing data filed within the Quality Management System and available onsite.