



## Certificate of Conformance

Certificate Number: EXMJ4340X33

Dated Issued: 03/28/2023

The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.<sup>1</sup>

<b>Product Item Number:</b>	X101835
<b>Item Description:</b>	ASK, CUTTING HEAD, 10PC
<b>Item Lot Number:</b>	X-0000000000033
<b>Expiration Date:</b>	2025-03

The blades were manufactured using a validated sterilization process:

<b>STERIS Gamma Process Run ID:</b>	1073-13455A		
<b>Processing Run Date:</b>	27-Mar-2023		
<b>Minimum Specified Dose (kGy):</b>	25.0	<b>Minimum Delivered Dose (kGy):</b>	27.5
<b>Maximum Specified Dose (kGy):</b>	45.0	<b>Maximum Delivered Dose (kGy):</b>	38.7

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company. Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE**.

  
 \_\_\_\_\_  
 Quality Department

  
 \_\_\_\_\_  
 Quality Manager

03/28/2023  
 \_\_\_\_\_  
 Date

3/29/23  
 \_\_\_\_\_  
 Date

<sup>1</sup> Manufacturing data filed within the Quality Management System and available onsite.