



Certificate of Conformance

Certificate Number: EXMXJ4355X39

Dated Issued: 04/26/2023

The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.¹

Product Item Number:	X101835
Item Description:	ASK, CUTTING HEAD, 10 PC
Item Lot Number:	X-0000000000039
Expiration Date:	2025-04

The blades were manufactured using a validated sterilization process:

STERIS Gamma Process Run ID:	1073-13810A		
Processing Run Date:	25-Apr-2023		
Minimum Specified Dose (kGy):	25.0	Minimum Delivered Dose (kGy):	28.6
Maximum Specified Dose (kGy):	45.0	Maximum Delivered Dose (kGy):	37.9


Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company. Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE**.




 Quality Department

4/26/23

 Date



 Quality Manager



 President

04-27-2023

 Date

¹ Manufacturing data filed within the Quality Management System and available onsite.