



Certificate of Conformance

Certificate Number: EXMXJ4358X144

Dated Issued: 06/13/2023

The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.¹

Product Item Number:	X100457
Item Description:	BLADE, BOX OF 5 AMALGATOME MD PKG
Item Lot Number:	X-0000000000144
Expiration Date:	2025/04

The blades were manufactured using a validated sterilization process:

STERIS Gamma Process Run ID:	1073-14406A		
Processing Run Date:	10-Jun-2023		
Minimum Specified Dose (kGy):	25.0	Minimum Delivered Dose (kGy):	27.0
Maximum Specified Dose (kGy):	45.0	Maximum Delivered Dose (kGy):	39.0

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company. Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE.**



 Quality Department



 Quality Manager

06/13/2023

 Date

6/29/23

 Date

¹ Manufacturing data filed within the Quality Management System and available onsite.