



## Certificate of Conformance

Certificate Number: EXMJ4557X58

Dated Issued: 12/28/2023

The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.<sup>1</sup>

|                             |                          |
|-----------------------------|--------------------------|
| <b>Product Item Number:</b> | X101835                  |
| <b>Item Description:</b>    | ASK, CUTTING HEAD, 10 PC |
| <b>Item Lot Number:</b>     | X-0000000000058          |
| <b>Expiration Date:</b>     | 2025-10                  |

The blades were manufactured using a validated sterilization process:

|                                      |             |                                      |      |
|--------------------------------------|-------------|--------------------------------------|------|
| <b>STERIS Gamma Process Run ID:</b>  | 1073-16896A |                                      |      |
| <b>Processing Run Date:</b>          | 26-Dec-2023 |                                      |      |
| <b>Minimum Specified Dose (kGy):</b> | 25.0        | <b>Minimum Delivered Dose (kGy):</b> | 27.4 |
| <b>Maximum Specified Dose (kGy):</b> | 45.0        | <b>Maximum Delivered Dose (kGy):</b> | 39.1 |

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company. Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE.**

  
 \_\_\_\_\_  
 Quality Department  
  
 \_\_\_\_\_  
 Quality Manager

12/28/23  
 \_\_\_\_\_  
 Date

12/28/2023  
 \_\_\_\_\_  
 Date

<sup>1</sup> Manufacturing data filed within the Quality Management System and available onsite.