



Certificate of Conformance

Certificate Number: EXMXJ4794X70

Dated Issued: 5/30/2024

The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.¹

Product Item Number:	X101835
Item Description:	ASK, CUTTING HEAD, 10 PC
Item Lot Number:	X-0000000000070
Expiration Date:	2026-05

The blades were manufactured using a validated sterilization process:

STERIS Gamma Process Run ID:	1073-18924A		
Processing Run Date:	28-Mar-2024		
Minimum Specified Dose (kGy):	25.0	Minimum Delivered Dose (kGy):	28.5
Maximum Specified Dose (kGy):	45.0	Maximum Delivered Dose (kGy):	38.2

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company.
Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE.**


Quality Department

5/30/24
Date


Quality Manager

05-30-2024
Date

¹ Manufacturing data filed within the Quality Management System and available onsite.