



Certificate of Conformance

Certificate Number: EXMXJ4812X75

Dated Issued: 7/9/2024

The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.¹

| | |
|-----------------------------|--------------------------|
| Product Item Number: | X101835 |
| Item Description: | ASK, CUTTING HEAD, 10 PC |
| Item Lot Number: | X-00000000000075 |
| Expiration Date: | 2026-05 |

The blades were manufactured using a validated sterilization process:

| | | | |
|--------------------------------------|-------------|--------------------------------------|------|
| STERIS Gamma Process Run ID: | 1073-19413A | | |
| Processing Run Date: | 07-Jul-2024 | | |
| Minimum Specified Dose (kGy): | 25.0 | Minimum Delivered Dose (kGy): | 28.6 |
| Maximum Specified Dose (kGy): | 45.0 | Maximum Delivered Dose (kGy): | 39.0 |

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company. Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE.**



 Quality Department



 Quality Manager

7/9/24

 Date

07-09-2024

 Date

¹ Manufacturing data filed within the Quality Management System and available onsite.