



Certificate of Conformance

Certificate Number: EXMXJ4963X87

Dated Issued: 09/30/2024

The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.¹

Product Item Number:	X101835
Item Description:	ASK, CUTTING HEAD, 10 PC
Item Lot Number:	X-00000000000087
Expiration Date:	2026-09

The blades were manufactured using a validated sterilization process:

STERIS Gamma Process Run ID:	1073-20486A		
Processing Run Date:	23-Sep-2024		
Minimum Specified Dose (kGy):	25.0	Minimum Delivered Dose (kGy):	28.7
Maximum Specified Dose (kGy):	45.0	Maximum Delivered Dose (kGy):	38.4

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company. Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE.**



Quality Department



Quality Manager

09/30/24

Date

9/30/24

Date

¹ Manufacturing data filed within the Quality Management System and available onsite.