



## Certificate of Conformance

Certificate Number: EXMXJ5038X94

Dated Issued: 11/25/2024

The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.<sup>1</sup>

<b>Product Item Number:</b>	X101835
<b>Item Description:</b>	ASK, CUTTING HEAD, 10 PC
<b>Item Lot Number:</b>	X-0000000000094
<b>Expiration Date:</b>	2026-10

The blades were manufactured using a validated sterilization process:

<b>STERIS Gamma Process Run ID:</b>	1073-21221A		
<b>Processing Run Date:</b>	11/22/2024		
<b>Minimum Specified Dose (kGy):</b>	25.0	<b>Minimum Delivered Dose (kGy):</b>	27.5
<b>Maximum Specified Dose (kGy):</b>	45.0	<b>Maximum Delivered Dose (kGy):</b>	39.1

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company. Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE**.



Quality Department



Quality Manager

11/25/2024

Date

11/25/2024

Date

<sup>1</sup> Manufacturing data filed within the Quality Management System and available onsite.