



Certificate of Conformance

Certificate Number: EXMXJ5062X99

Dated Issued: 12/18/2024

The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.¹

Product Item Number:	X101835
Item Description:	ASK, CUTTING HEAD, 10 PC
Item Lot Number:	X-0000000000099
Expiration Date:	2026-11

The blades were manufactured using a validated sterilization process:

STERIS Gamma Process Run ID:	1073-21586A		
Processing Run Date:	12-16-2024		
Minimum Specified Dose (kGy):	25.0	Minimum Delivered Dose (kGy):	29.9
Maximum Specified Dose (kGy):	45.0	Maximum Delivered Dose (kGy):	39.1

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company. Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE.**



Quality Department



Quality Manager

12/18/2024

Date

12/19/2024

Date

¹ Manufacturing data filed within the Quality Management System and available onsite.