



Certificate of Conformance

Certificate Number: EXMJ5165X56

Dated Issued: 01/20/2025

The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.¹

Product Item Number:	X101789
Item Description:	PACK, AMALGATOME SD 2" BLADE, 5 PC
Item Lot Number:	X-0000000000056
Expiration Date:	01/03/2027

The blades were manufactured using a validated sterilization process:

STERIS Gamma Process Run ID:	1073-21934A		
Processing Run Date:	01/17/2025		
Minimum Specified Dose (kGy):	25.0	Minimum Delivered Dose (kGy):	28.3
Maximum Specified Dose (kGy):	45.0	Maximum Delivered Dose (kGy):	39.1

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company. Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE**.

Rebecca Coulee
Quality Department

[Signature]
Quality Manager

01/20/25
Date

2/20/25
Date

¹ Manufacturing data filed within the Quality Management System and available onsite.