



Certificate of Conformance

Certificate Number: EXMXJ5279X57

Dated Issued: 03/19/2025

The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.¹

Product Item Number:	X101789
Item Description:	PACK, AMALGATOME SD 2" BLADE, 5 PC
Item Lot Number:	X-0000000000057
Expiration Date:	03/05/2027

The blades were manufactured using a validated sterilization process:

STERIS Gamma Process Run ID:	1073-22774A		
Processing Run Date:	03/18/2025		
Minimum Specified Dose (kGy):	25.0	Minimum Delivered Dose (kGy):	28.8
Maximum Specified Dose (kGy):	45.0	Maximum Delivered Dose (kGy):	39.0

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company. Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE**.



 Quality Department



 Quality Manager

03/19/25

 Date

3/19/25

 Date

¹ Manufacturing data filed within the Quality Management System and available onsite.