



Certificate of Conformance

Certificate Number: EXMXJ5309X59

Dated Issued: 04/16/2025

The blades were produced in accordance with Exsurco current design specification using only approved manufacturing processes and conforms to all established quality inspection and test criteria.¹

Product Item Number:	X101789
Item Description:	PACK, AMALGATOME SD 2" BLADE, 5 PC
Item Lot Number:	X-000000000059
Expiration Date:	03/26/2027

The blades were manufactured using a validated sterilization process:

STERIS Gamma Process Run ID:	1073-23082A		
Processing Run Date:	04/14/2025		
Minimum Specified Dose (kGy):	25.0	Minimum Delivered Dose (kGy):	29.7
Maximum Specified Dose (kGy):	45.0	Maximum Delivered Dose (kGy):	38.9

Exsurco Medical is an ISO 13485:2016 certified and an FDA registered company.
Exsurco Medical is committed to a policy of **QUALITY WITHOUT COMPROMISE**.

Release Cover

Quality Department

Tony E. Tamm, Jr.

Quality Manager

04/14/25

Date

4/16/25

Date

¹ Manufacturing data filed within the Quality Management System and available onsite.